

Instructions

1 Collect Specimen.

Verify Temperature

2 Peel Back Label.

3 Read the Results.

Read results at five (5) minutes.
Results stable up to sixty (60) minutes

NEGATIVE

NON-NEGATIVE

INVALID

	Abnormal	Normal
Oxidant	 Low High	 Negative
S.G.	 1.000 Low 1.035 High	 1.003 1.030
pH	 2.0 3.0 10.0 11.0 12.0	 4.0 9.0

SAMPLE ONLY - Color chart NOT Exact